

1 Supplemental Information for:

2 Standardization of a UV LED Peak Wavelength,
3 Emission Spectrum, and Irradiance Measurement
4 and Comparison Protocol

5 *Kari Sholtes, Ryan Keliher, and Karl G. Linden**

6 **Corresponding author: Karl.Linden@colorado.edu*

7 Department of Civil, Environmental, and Architectural Engineering, University of Colorado

8 Boulder, Boulder CO 80309 USA

9

Introduction

This supplemental information (SI) contains the standardized protocol for both spectrometric and radiometric measurements of UV LEDs. In an effort to make the protocol more practical, it contains quality control and quality assurance steps which serve as a checklist to ensure all steps and comparisons were completed. The standard protocol provides guidelines to ensure that measurements are valid, particularly determining the distances for measuring the emission spectra and irradiance. For example, the 15 and 20 cm distances used in the study informing the protocol were determined by ensuring the peak wavelength as measured with a spectrometer was less than 85% of the saturation value for all UV LEDs. These values initially corresponded with the conventional wisdom regarding irradiance measured at a distance, but the 15 cm distance was not sufficient to meet new standards for irradiance measurements. The protocol details that all irradiance measurement need to be taken at $>7\times$ the largest dimension of the UV LED array, but it is not prescriptive about what the numeric distances should be. Also contained in this SI are general instructions for how to use the protocol and suggested documentation.

Quality Assurance and Quality Control (QA/QC) Spectrometer and Radiometer Measurements

In order to establish QA/QC for radiometer and spectrometer measurements, the following steps are recommended for each type of equipment and associated measurement:

Spectrometer QA/QC

- 1) Verify that the spectrometer (with either spectrometric or spectroradiometric calibration) has been calibrated over the applicable range by a qualified or accredited third party or using a standard method within one year

- 33 2) Test a UV plasma source with a well-defined wavelength emission profile (i.e., 253.7 nm low
34 pressure mercury arc) near the range of the target UV LED wavelengths to be tested
- 35 a) Verify ± 1 nm difference from NIST atomic spectra database (“NIST: Atomic Spectra
36 Database Lines Form” n.d.) expected value
- 37 3) Follow the protocol found in the SI, Emission Spectrum and Irradiance Measurement Protocol
38 for UV LEDs
- 39 4) Take all measurements in duplicate
- 40 a) Ensure measurements are within the reported uncertainty due to the calibration range as
41 indicated on the calibration certificate
- 42 5) Repeat all measurements at a minimum of two different heights
- 43 a) Ensure repeated measurements are within the reported uncertainty based on the calibration
44 range
- 45 6a) Compare spectrometer results to UV LED manufacturer or device developer’s source spectrum
46 data
- 47 a) Calculate the mean wavelength of the replicate samples and the manufacturer’s data
- 48 b) Verify all measurements are ± 1 nm difference from the mean
- 49 OR
- 50 6b) Test with secondary spectrometer and compare with primary results (annual)
- 51 a) Calculate the mean wavelength of the replicate samples from both spectrometers

b) Verify all measurements are ± 1 nm difference from the mean calculated above

Radiometer QA/QC

1) Ensure that the radiometer is calibrated over the ultraviolet range of interest (typically every 2 nm) to accurately characterize the target UV LED spectrum (as opposed to calibration at a single wavelength) and verify that the sensor has a linear response (meaning that the error is uniform) across the applicable wavelength range

2) Verify that the radiometer and detector have been calibrated over the applicable range by an accredited third party using standards traceable to a national laboratory such as the National Institute of Standards and Technology (NIST) within one year

3) Follow the protocol found in SI, Emission Spectrum and Irradiance Measurement Protocol for UV LEDs

4) Take all measurements in duplicate

a) Ensure repeat measurements are within the reported error or uncertainty as published on the calibration certificate by the radiometer manufacturer

5) Repeat all measurements at a minimum of two different distances between the emission source and the sensor plane of calibration

a) Compare the radiometric data for various heights using the calculated radiant intensity from the inverse square law:

Equation S1: $E_1 d_1^2 = E_2 d_2^2 = I$

where E is the irradiance (Watts/cm²), d is the distance (cm), and I is the radiant intensity (Watts/sr).

b) Calculate the mean of the radiant intensity, I, for all matched radiometric samples (various heights and duplicates)

c) Calculate the percent error from the mean for each measurement

Equation S2: Percent error from the mean = $\frac{(\lambda_{experimental} - \lambda_{expected})}{\lambda_{expected}} * 100$

d) All radiant intensities should vary by the less than the error as reported on the calibration certificate or $\pm 5\%$ error from the mean (whichever is larger) across various heights and duplicates using the same equipment operated under the same electrical conditions

6a) If possible, compare radiometer results to UV LED manufacturer or device developer's irradiance data (when new or otherwise comparable)

a) Calculate the mean of the radiant intensity, I, for all radiometric samples (various heights and duplicates)

b) Verify that radiant intensities vary by $\pm 10\%$ error from the mean

OR

6b) Test with secondary radiometer and compare with primary results (when new or otherwise comparable)

a) Calculate the mean of the radiant intensity, I, for all radiometric samples (various heights and duplicates)

b) Verify that radiant intensities vary by $\pm 10\%$ error from the mean

UV LED Source Emission Spectrum and Irradiance Measurement Protocol

General Instructions

1) Perform protocol testing in an area where stray light < 400 nm can be minimized

2) Take photos of the setup for reporting purposes

3) Use proper Personal Protective Equipment

a) Properly fitted UV protective glasses and gloves

b) Avoid unprotected skin

c) No access to work area for those not protected by goggles and proper clothing

4) Verify that all equipment is in good condition prior to starting the protocol

a) Tampering with measurement equipment can void calibrations and negate all measurements

5) Verify that all equipment has been calibrated over the applicable range within one year

a) Spectrometers need to be calibrated by a qualified or accredited third party or using a standard method

b) Radiometers with a detector traceable to a national laboratory such as NIST needs to be calibrated over the applicable range by an accredited third party

6) Identify the UV-C LED emission plane

- 109 a) The emission plane is the chip itself if no lens is present
- 110 b) The emission plan is the face of the lens if a lens is present
- 111 c) Mark the UV-C LED emission plane for repeatable measurements (if applicable)
- 112 7) Measure and document the largest dimension of the UV LED source
- 113 a) If one UV LED, measure the largest diagonal dimension of the chip
- 114 b) If multiple UV LEDs or chips in an array, measure the largest diagonal distance across all
- 115 UV LEDs
- 116 c) Calculate 7x this largest dimension of the UV LED (s), this distance will mark the closest
- 117 measurement distances allowable for the radiometric measurements

118 Spectrometer Protocol Initial Documentation

- 119 1) Document the spectrometer manufacturer, name/type of spectrometer and sensor
- 120 a) Document the size and type of integrating sphere (if applicable)
- 121 2) Document the temperature in the facility and time/location of testing
- 122 3) Ensure spectrometer is calibrated to the proper wavelength range
- 123 a) UV: 200-400 nm is recommended to minimize uncertainty error associated with the
- 124 calibration range
- 125 -OR-
- 126 UV/Vis: 200-1000 nm if required, but the larger wavelength range has a larger uncertainty
- 127 associated with it

- 128 b) Document the calibration type (spectroradiometric or spectrometric), date, and range
- 129 c) Document the error associated with the calibration range (per manufacturer)
- 130 4) If a lens was present during calibration, verify that the lens is present, intact, and the calibration
- 131 is valid
- 132 a) Document the surface area of your sensor and the type of lens if applicable
- 133 5) Ensure usage of proper settings for integration time, nonlinearity correction, scans to average,
- 134 etc., depending on the operation of your UV LEDs and your data collection and
- 135 experimentation goals
- 136 a) Read the spectrometer manual to understand operation and proper settings
- 137 b) Call the spectrometer manufacturer to discuss testing goals and suggested input values for
- 138 the integration time, corrections, etc., if needed
- 139 c) Integration time can be increased to improve the signal to noise ratio by the square root of
- 140 the integration time, but this will increase the total acquisition time
- 141 i) Ensure no pixels are saturated (limit maximum saturation to 85% of saturation)
- 142 d) Spatial averaging is not recommended for UV LED spectra due to the flattening of the peak
- 143 i) Applicable with a spectrum that is relatively flat
- 144 e) Scans to average can be increased to increase the signal to noise ratio
- 145 i) Total acquisition time is a product of the scans to average times the integration time


- 146 f) Boxcars (0-5) can be increased to increase the signal to noise ratio but it will also flatten
- 147 the peaks
- 148 g) Nonlinearity correction should be used to correct for the non-linear response to sensor or
- 149 cosine correctors
- 150 h) Signal to noise ratio simply refers to the true signal of a spectrum relative to the unwanted
- 151 signal (from numerous sources) in a spectrum
- 152 i) Document all settings with an explanation/justification of their use
- 153 6) Ensure access to current and valid digital calibration data if using a spectrometer with
- 154 radiometric calibration before testing

155 Spectrometer Protocol Experimental Set-up and Measurements

- 156 1) Locate the sensor plane on the spectrometer sensor from which to take measurements
- 157 a) Mark the sensor plane if necessary for repeatable measurements
- 158 2) Adjust distance from the UV-C LED emission plane and the sensor plane to the experimental
- 159 distance (s)
- 160 a) If using a spectrometer with a radiometric calibration, ensure that the following conditions
- 161 are met:
- 162 i) the sensor is normal to the beam of light emitted from the UV source
- 163 ii) the sensor is located at the centerline of the UV source

- 164 iii) the sensor is located at a distance of greater than 7x the largest dimension of the UV
165 LED light source
- 166 b) If using a spectrometer with only a spectrometric calibration, simply ensure that you are
167 capturing your light source without saturating the spectrometer
- 168 c) If using an integrating sphere, follow integrating sphere instructions for set-up
- 169 3) Plug in and turn on the spectrometer and any associated equipment (i.e. laptops, etc.)
- 170 4) Store dark spectrum data according to manufacturer's instructions
- 171 a) Typically, the cap will be kept on for this reading then can be removed for UV
172 measurements
- 173 5) Ensure that minimal measurable light < 400 nm is present in the vicinity (with the UV source
174 OFF)
- 175 6) Take one ambient/background reading with the UV source OFF in the experimental set-up
- 176 a) Save these data
- 177 7) Ensure PPE are adjusted properly
- 178 a) UV protective goggles fit properly
- 179 b) Skin which could be in contact with UV light is covered
- 180 c) No access to work area for those not protected by goggles and proper clothing
- 181 8) Turn UV source ON

- 182 a) Collect UV spectrum at 5 and 10 minutes
- 183 b) Save data
- 184 9) Turn UV source OFF
- 185 10) Analyze the spectral data for the 5 and 10 minute time points:
- 186 a) Visualize the data and remove any negative values or anomalous data
- 187 b) Identify and document the maximum value of the spectrum
- 188 i) If data are noisy near the peak, a data smoothing technique may be required (i.e.,
- 189 splining)
- 190 c) Normalize all values to the peak value
- 191 d) Identify the thresholds on either tail where the curve nears approximately 0.5% of the value
- 192 of the peak; this will serve as the data analysis cut off point (see Figure SI 1).
- 193 i) If the curve does not approach zero on both sides of the peak, first analyze the
- 194 ambient/background light.
- 195 (1) Subtract the ambient/background light from the raw curve data to see if the curve
- 196 approaches zero
- 197 (2) Try zeroing the device again or call the spectrometer manufacturer for
- 198 troubleshooting
- 199 e) Find the area under the curve between the two 0.5% peak value thresholds of each tail
- 200 f) Document the wavelength range of curve between the 0.5% peak value thresholds



- 201
- 202 Figure **Error! No sequence specified.**: Normalized Emission Spectrum Area under the Curve
- 203 (AUC) Quantification with Thresholds
- 204 g) Compare the visualized data, the peak, the range, and the area under the curve between the
- 205 5 and 10 min time points.
- 206 i) Calculate the mean of the peak wavelength and the area under the curve
- 207 ii) Calculate the percent error from the mean for the peak and the area under the curve for
- 208 each measurement (Equation S2)
- 209 h) Verify the peak wavelengths vary by less than the reported spectrometer uncertainty (this
- 210 is a function of the calibration range)
- 211 i) If radiometrically calibrated: verify that the areas under the curve at the same height are
- 212 within $\pm 5\%$ error from the mean, if applicable

- 213 j) If radiometrically calibrated: verify that the normalized areas under the curve are within
214 $\pm 1\%$ error from the mean
- 215 11) Adjust the UV emission plane and the spectrometer sensor plane to the second experimental
216 distance
- 217 a) If there is no second experimental distance planned, simply develop an arbitrary one for
218 QA/QC purposes
- 219 b) If using an integrating sphere, simply run the experiment again after removing all sources
220 from the integrating sphere for duplication purposes
- 221 c) If using a spectrometer with a radiometric calibration, ensure that the following conditions
222 are met:
- 223 i) the sensor is normal to the beam of light emitted from the UV source
- 224 ii) the sensor is located at the centerline of the UV source
- 225 iii) the sensor is located at a distance of greater than 7x the largest dimension of the UV
226 LED light source
- 227 d) If using a spectrometer with only a spectrometric calibration, simply ensure that you are
228 capturing your light source without saturating the spectrometer
- 229 12) Ensure that minimal measurable light < 400 nm is present in the vicinity (with the UV source
230 OFF)
- 231 13) Take one ambient/background reading with the UV source OFF in the experimental set-up

232 a) Save these data

233 14) Ensure PPE are adjusted properly

234 a) UV protective goggles fit properly

235 b) Skin is covered which could be in contact with UV light

236 c) No access to work area for those not protected by goggles and proper clothing

237 15) Turn UV source ON

238 a) Collect UV spectrum at 5 or 10 minutes

239 b) Save data

240 16) Turn UV source OFF

241 17) Replace the cap to protect the sensor on the spectrometer

242 18) Repeat all of the above steps again

243 a) Save files as Trial 2 of each test

244 19) Analyze and compare the spectral data as described in Step 10 above

245 20) Verify that the peak wavelengths vary by less than the reported spectrometer uncertainty

246 21) If radiometrically calibrated: verify that the areas under the curve at the same height are within

247 $\pm 5\%$ error from the mean, if applicable

248 22) If radiometrically calibrated: verify the normalized areas under the curve at all heights are

249 within $\pm 1\%$ error from the mean, if applicable

- 250 23) Compare these data to either the manufacturer's data or to the same analysis using a different
251 spectrometer
- 252 a) Verify all peak wavelength data are within ± 1 nm of the mean (even when produced by
253 different spectrometers at different heights)
- 254 b) If radiometrically calibrated: verify that the area under the curve for each height are within
255 10% error from the mean and from radiometer irradiance data at the same height, if
256 applicable
- 257 c) If radiometrically calibrated: verify all normalized areas under the curve are within $\pm 5\%$ error
258 from the mean, if applicable

259

260 Radiometer Protocol Initial Documentation

- 261 1) Document the radiometer manufacturer, name/type of radiometer and sensor
- 262 2) Document the temperature in the facility and time/location of testing
- 263 3) Ensure radiometer and sensor is calibrated to the proper wavelength range and within
264 calibration date by an accredited third party
- 265 a) Document the calibration date and range
- 266 b) Document the error associated with the calibration range
- 267 4) Avoid direct irradiation of the radiometer sensor entrance slit with either an integrating sphere
268 or diffuser.

- 269 5) Ensure integrating sphere or diffuser lens is present (if applicable), intact, and the calibration
270 is valid (i.e., that the sensor lens has not been removed or damaged)
- 271 a) Document the surface area of sensor and the type of lens/diffuser (if applicable)
- 272 b) Document the size and type of integrating sphere (if applicable)
- 273 6) Ensure usage of proper radiometer and sensor depending on the operation of your UV LEDs
274 and your data collection and experimentation goals
- 275 a) Call the radiometer manufacturer to discuss testing goals if needed
- 276 7) Ensure access to current and valid digital calibration data before testing

277 Radiometer Protocol Experimental Set-up and Measurements

- 278 1) Locate the sensor plane on the radiometer sensor from which to take your measurements
- 279 a) This information is likely available in the manual
- 280 b) Mark the sensor plane if necessary for repeatable measurements
- 281 2) Adjust and document the distance from the UV-C LED emission plane and the sensor plane to
282 the experimental distance (s)
- 283 a) Ensure that the sensor is normal to the beam of light emitted from the UV source
- 284 b) Ensure that the radiometer sensor is located at the centerline of the UV source
- 285 c) Ensure that the sensor plane is located at a distance of greater than 7x the largest dimension
286 of the UV LED light source
- 287 d) Or follow the integrating sphere instructions (as applicable)

- 288 3) Plug in and turn on the radiometer and any associated equipment (i.e. laptops, etc.)
- 289 4) Ensure that minimal measurable light < 400 nm is present in the vicinity (with the UV source
290 OFF)
- 291 a) Measurable light < 400 nm ideally should be less than 0.1% of the irradiance measurement
292 at the peak wavelength
- 293 5) Use the average peak wavelength for the applicable UV-C LED (s) from the spectrometer
294 analysis for all radiometric analysis
- 295 a) Determine the calibration factor associated with the peak wavelength
- 296 b) Apply that calibration factor to all applicable radiometric measurements – refer to the
297 radiometer manual
- 298 6) Zero the device with a cover over the detector once the calibration factor is set
- 299 7) Take one ambient/background reading with the UV source OFF in the experimental set-up
- 300 a) Save these data
- 301 8) Ensure PPE are adjusted properly
- 302 a) UV protective goggles fit properly
- 303 b) Skin which could be in contact with UV light is covered
- 304 c) No access to work area for those not protected by goggles and proper clothing
- 305 9) Turn the radiometer data logger on (if applicable)

- 306 10) Turn UV source ON
- 307 a) Collect radiometric data continually with a data logger or every 30 seconds until steady
- 308 state has been achieved or the irradiance is changing by less than 1% over the time horizon
- 309 of the planned experiments (if applicable)
- 310 b) Record the time frame to achieve steady state
- 311 c) Continue to take readings for another 5 minutes (in the case of a data logger) or take another
- 312 reading in 5 minutes
- 313 d) Save or record data
- 314 11) Turn UV source OFF
- 315 12) Calculate the error from the mean for the steady state experiments
- 316 a) Verify that the steady state measurements are all within the radiometer uncertainty (if
- 317 published) or $\pm 1\%$ error from the mean
- 318 13) Adjust and document the second experimental distance from the UV emission plane to the
- 319 radiometer sensor plane
- 320 a) If there is no second experimental distance, simply develop an arbitrary one for QA/QC
- 321 purposes
- 322 b) If using an integrating sphere, simply run the experiment again after removing and
- 323 replacing the UV source
- 324 c) Ensure that the sensor is normal to the beam of light emitted from the UV source

- 325 d) Ensure that the radiometer sensor is located at the centerline of the UV source
- 326 e) Ensure that the sensor plane is located at a distance of greater than 7x the largest dimension
- 327 of the UV LED light source
- 328 14) Ensure PPE are adjusted properly
- 329 a) UV protective goggles fit properly
- 330 b) Skin which could be in contact with UV light is covered
- 331 c) No access to work area for those not protected by goggles and proper clothing
- 332 15) Turn the radiometer data logger on (if applicable)
- 333 16) Turn UV source ON
- 334 a) Collect radiometric data until steady state has been achieved or the irradiance is changing
- 335 by less than 1% over the time horizon the planned experiments (if applicable)
- 336 b) Record the time frame to achieve steady state
- 337 c) Continue to take readings for another 5 minutes (in the case of a data logger) or take another
- 338 reading in 5 minutes
- 339 d) Save or record data
- 340 17) Turn UV source OFF
- 341 18) Replace the cap to protect the diffuser lens on the radiometer

342 19) Compare the radiometric data for various heights using the calculated radiant intensity from
343 the inverse square law:

344 a) The inverse square law of electromagnetic radiation explains the relationship between a
345 source and measurement distance. Irradiance is proportional to the inverse of the squared
346 distance (d in cm): $E = \frac{I}{d^2}$ where E is the irradiance (Watts/cm²) and I is the radiant
347 intensity (Watts/sr). In measuring the irradiance at two distances, the equation can be
348 manipulated to Equation S1

349 b) Calculate the mean of the radiant intensity, I, for all radiometric samples for each UV LED

350 c) Calculate the percent error from the mean for each measurement

351 d) The radiant intensity should vary by $\pm 5\%$ error from the mean across various heights and
352 duplicates

353 e) If data are not comparable

354 i) Ensure no ambient/background irradiance

355 ii) Retest with enhanced attention to the time required to achieve steady-state emission,
356 measurement distances, testing along the centerlines of the UV LEDs and sensor, and
357 ensuring that the sensor face is parallel to the face of the UV LEDs

358 iii) Consider irregular angular emission and perhaps increase the distance between
359 emission and sensor planes

360 20) Determine the average time to achieve steady state for these UV-C LEDs, and use this for
361 future radiometric and/or spectrometric experiments

362 a) If this time to achieve steady state is significantly different from the time-frames tested in
363 the spectrometer measurement section, go back to re-test the spectrometer to ensure that
364 the emission spectrum does not change over time (especially if using a spectrometer with
365 a radiometric calibration)

366 21) Repeat all of the above steps again for the pre-determined steady-state times only

367 a) Save files as Trial 2 of each test

368 22) Analyze and compare the irradiance data as described above

369 a) Ensure the steady-state irradiance of all duplicate UV-C LED data (same heights, same
370 UV-C LEDs, same radiometer) are within the reported radiometer uncertainty or within
371 $\pm 1\%$ distance from the mean

372 b) Ensure the steady-state irradiance of the same UV-C LEDs over various heights (but the
373 same radiometer) are comparable: the radiant intensity (irradiance times the squared
374 distance from the source emission face to the sensor plane) should vary by $\pm 5\%$ error from
375 the mean across various heights and duplicates

376 23) Compare these data to either the manufacturer's data or to the same analysis using different
377 radiometric or spectroradiometric data

378 a) Find the radiant intensity for all irradiance data

379 b) Average the radiant intensity for each UV-C LED at both (all) experimental heights and
380 ensure that all measurements are within $\pm 10\%$ error from the mean

381 QA/QC Verification Summary

382 Spectrometer

- 383 • Data for one spectrometer, various heights and UV LED operation times
 - 384 ○ Verify that the average peak wavelengths vary by less than the reported
 - 385 spectrometer uncertainty
 - 386 ○ If radiometrically calibrated: verify that the areas under the curve at the same height
 - 387 are within $\pm 5\%$ error from the mean, if applicable
 - 388 ○ If radiometrically calibrated, verify that all normalized emission spectrum area
 - 389 under the curve varies by $\pm 1\%$ error from the mean

390 • Data for more than one spectrometer, various heights and UV LED operation times

- 391 ○ Verify that the average peak wavelengths vary by ± 1 nm from the average peak
- 392 wavelength
- 393 ○ If radiometrically calibrated: verify that the areas under the curve at the same height
- 394 are within $\pm 10\%$ error from the mean, if applicable
- 395 ○ If radiometrically calibrated: verify that the normalized emission spectrum area
- 396 under the curves vary by $\pm 5\%$ error from the mean (can compare different heights
- 397 together with normalized AUC)

398 Radiometer

- 399 • Data for one radiometer, one height, two steady-state UV LED operation times

- 400 ○ Verify that the steady state irradiances vary by less than the reported radiometer
- 401 uncertainty or $\pm 1\%$ error from the mean if no uncertainty or error is reported by the
- 402 manufacturer

- 403 • Data for one radiometer, various heights and steady-state UV LED operation times

- 404 ○ Verify that the radiant intensities vary by $\pm 5\%$ error from the mean

- 405 • Data for more than one radiometer (or spectrometer with radiometric calibration), various
- 406 heights and steady-state UV LED operation times

- 407 ○ Verify that the radiant intensities for all heights vary by $\pm 10\%$ error from the mean